

# NordVal International Certificate

Issued for:	Compact Dry EC Method for the Enumeration of <i>Escherichia coli</i> and coliforms
NordVal No:	036
First approval date:	01 December 2008
Renewal date:	22 November 2022
Valid until:	01 December 2024

## Compact Dry EC

### Manufactured and supplied by:

Nissui Pharmaceutical Co.Ltd,  
3-23-9 Ueno,  
Taito-ku, Tokyo, 110-8736  
Japan

The performance of this has been compared to the reference methods:

- ISO 16649-2:2001: "Microbiology of food and animal feeding stuffs. Horizontal method for the enumeration of beta-glucuronidase-positive *Escherichia coli*. Part 2: Colony-count technique at 44 degrees C using 5-bromo-4-chloro-3-indolyl beta-D-glucuronide."


and

- ISO 4832:2006: "Microbiology of food and animal feeding stuffs. Horizontal method for the enumeration of coliforms. Colony-count technique."

The validation studies have been conducted by Campden, UK, according to t ISO 16140-2:2016, and concludes Compact Dry EC provide equivalent results to the reference methods.


The production of Compact Dry EC is certified according to ISO 9001 and ISO 13485.

Yours sincerely,



Hilde Skår Norli  
Chair of NordVal International

Date: 22 November 2022



Eystein Oveland  
NMKL Executive Director

## PRINCIPLE OF THE METHOD

Compact Dry EC is a ready-to-use dry chromogenic plate for enumeration of *E.coli* and coliforms. Pre-treat the samples according to ISO 6687 or NMKL 91. The medium contains two kinds of chromogenic enzyme substrates: Magenta-Gal and X-Gluc. *E.coli* forms blue colonies. The total coliform group count is the sum of both the red and blue colonies.

An aliquot of 1 ml of an appropriate dilution is plated onto Compact Dry EC plate. The incubation conditions tested in the study were  $37 \pm 1^\circ\text{C}$  for  $24 \pm 2\text{h}$ .

## FIELD OF APPLICATION

The method has been tested on enumeration of *Escherichia coli* and coliforms in a broad range of foods.

## HISTORY

In 2007, the method was validated according to the ISO 16140:2003. Every two years until 2018 the method has been renewed without any additional studies.

In 2018 a renewal study was performed to comply with the requirements for relative trueness and accuracy profile in the new standard ISO 16140-2:2016. As the design of the Inter-laboratory study (ILS) is the same for the 2003 and 2016 versions of ISO16140, the data from the ILS data of 2007 are re-evaluated using the new statistical approach outlined in ISO16140-2:2016.

## METHOD COMPARISON STUDY

### Relative trueness study

The trueness study is a comparative study between results obtained by the reference method and the results of the alternative method. Different categories, types and items were tested as shown in Table 1 below.


*Table 1. Categories and types tested*

Category	Types	No. of samples
Milk and dairy products	Dry milk product e.g. milk powder, powder for milk based desserts, dried infant formula	5
	Dairy products e.g. ice-cream, yogurts, cream, hard cheese, soft cheese, raw milk cheese	5
	Pasteurised milk products e.g. skimmed, semi-skimmed, full fat and flavoured milks	5
Fresh produce and fruits	Cut ready to eat fruit e.g. fruit mixes, fruit juices	5
	Cut ready to eat vegetables e.g. Bagged pre-cut salads and shredded carrot, cabbage, vegetable juices	5
	Leafy greens/Sprouts e.g. soy, mung, alfalfa,	5
Raw poultry and meats (Combined category raw/ RTC meats and poultry)	Fresh poultry cuts e.g. turkey breast, turkey fillet	5
	Fresh mince e.g. lamb, beef, pork	5
	Processed ready to cook e.g. frozen patties, marinated kebabs, seasoned chicken breasts	5
Ready to eat foods (Combined category RTE/RTRH meats and poultry)	Ready to eat poultry e.g. turkey fillet, chicken sausage, pate	5
	Cooked fish products e.g. prawns, terrine, pate, smoked fish	5
	Cooked meat e.g. ham, salami, pate, corned beef	5
Multi component foods or meal components	Ready to re-heat refrigerated food e.g. cooked chilled foods, rice and pasta, products	5
	Ready to re-heat food frozen e.g. fries, pizza	5
	Composite foods with substantial raw ingredients e.g. pasta salads, sandwiches, deli-salads	5


The relative trueness is illustrated by the use of a Bland-Altman plot, i.e. the difference (bias) between paired samples analysed with the reference method and the alternative method respectively, plotted against the mean values obtained by the reference method. In the plot, Upper and Lower limits are included as the bias  $\pm 2$  times the standard deviation of the bias. The Bland-Altman Plot in Figure 1, illustrates the difference obtained in the enumeration of *E.coli* and total count by the alternative and the reference method, respectively.

Figure 1. Bland-Altman Plot of the enumeration of *E.coli* and total count in foods

*E. coli*:



Coliforms:



It is expected that no more than 1 in 20 data values will lie outside the 95% confidence levels (upper limit and lower limits).

### For E.coli:

For 'All Categories' there are five in 75 values which lie outside the CLs. This is a little more than the expectation of less than one in 20. There was no identifiable trend in the data outside the CLs, which covered 4 different food categories, 4 different inoculated strains and 3 different seeding/spiking protocols.

### For coliforms:

For 'All Categories' there are six in 75 values which lie outside the CLs. This is a little more than the expectation of less than one in 20. Of the six points outside of the CLs, the data covered 3 different food categories, and 3 different inoculated strains.

### Accuracy profile

The accuracy profile study is a comparative study between the results obtained by the reference method and the results of the alternative method. Each item used were artificially contaminated obtaining three target levels; low ( $10^2$  cfu/g), medium ( $10^4$  cfu/g) and high ( $10^6$  cfu/g). Five test portions of each level of each item were analysed, resulting in 150 samples. The tested categories, types, items and inoculated strains are provided in the Table 2.


Table 2. Categories, types and food items

Category	Types	Strains – E.coli	Strains - Coliforms	Items
Dairy products	Pasteurised dairy products	<i>E. coli</i> CRA 1476 from dried milk	<i>E. coli</i> CRA 1476 from dried milk	Pasteurised cream
		<i>E. coli</i> NCTC 8008	<i>Enterobacter agglomerans</i> CRA 5613 from milk powder	Cream cheese
Fruits and vegetables	Fresh produce	<i>E. coli</i> ATCC 25922	<i>E. hermanii</i> CRA 7477 from sesame seeds	Ready to cook Vegetable preparation
		<i>E. coli</i> NCIMB 700555	<i>Citrobacter amalonaticus</i> CRA 7458 from beansprouts	Vegetable juice
Raw poultry and meats (Combined category raw/ RTC meats and poultry)	Fresh meat	<i>E. coli</i> CRA 16041 from raw ground mince	<i>Enterobacter aerogenes</i> NCTC 10006	Pork mince
		<i>E. coli</i> CRA 1593 from poultry	<i>Citrobacter freundii</i> NCTC 9750	Raw bacon
Ready to eat foods (Combined category RTE/RTRH meats and poultry)	Cooked fish products e.g. prawns	<i>E. coli</i> CRA 2003 isolated from fish	<i>E. coli</i> CRA 2003 from fish	
		<i>E. coli</i> CRA 1968 isolated from lamb	<i>Klebsiella oxytoca</i> ATCC 15926	Fish pate
Multi component foods	Composite foods with raw ingredients	<i>E. coli</i> CRA 16044 isolated from beef	<i>Enterobacter agglomerans</i> CRA 5513 from skimmed milk powder	Sandwiches
		<i>E. coli</i> CRA 1265 dried foods	<i>E. adecarboxylata</i> CRA 5501 from skimmed milk powder	Cooked chilled rice

The total number of samples analysed for both *E.coli* and Coliforms with both methods were 150. The statistical results and the accuracy profiles are provided in the Figures 2 to 6.

Figure 2. Dairy products

*E.Coli*:



Coliforms:





Figure 3. Fruit and vegetable products

E.Coli:




Coliforms:



Figure 4. Meat and poultry

E.Coli:



Coliforms:





Figure 5. Ready to eat foods

E.Coli:



Coliforms:








Figure 6. Multi component foods

E.Coli:



Coliforms:





The observed profiles are within the 0.5 log AL or the recalculated AL limit calculated according to ISO16140-2:2015 section 6.1.3.3.

For both *E.coli* and coliforms, the accuracy profiles fulfil the performance criteria after the permitted recalculation and the alternative method is accepted as being equivalent to the reference methods.

## Selectivity (inclusivity/ exclusivity)

Inclusivity is the ability of an alternative method to detect the target analyte from a wide range of strains.

### *For E.coli:*

In the original study: 31 strains were studied. All 31 strains grew and produced typical colonies on the Compact Dry EC medium. By comparison, 5 strains failed to grow in the TBX medium (ISO16649-2:2001) and one strain yielded atypical colonies.

In the renewal study of 2018: Of the 20 inclusivity strains tested all strains were detected using both the alternative and reference method.

### *For coliforms:*

In the original study: All 33 coliform strains produced typical colonies in VRBA (ISO 4832) and Compact Dry EC medium.

In the renewal study from 2018: Of the 20 inclusivity strains tested 18 strains were detected using the alternative and the reference methods. Those not detected by either method were *Shimwellia blattae* NCTC 12127 and *Klebsiella rhinoscleromatis* CRA 4272.

Exclusivity is the lack of interference from a relevant range of non-target strains of the alternative method.

### *For E.coli:*

In the original study: The results from the 21 strains of non-target organisms showed that the majority (19 cultures) failed to grow or produced atypical colonies by both methods. Two strains of *Shigella* did yield typical colonies by both methods, which is not surprising because strains of *Shigella* have  $\beta$ -glucuronidase activity which would give rise to typical colonies with chromogenic media developed to show this activity.

In the current study (2018): Of the 10 exclusivity strains tested, none were detected by either the alternate or reference methods.

### *For coliforms:*

In the original study: The results from the 20 strains of non-target organisms used to determine the exclusivity of the EC method showed that 9 strains did not grow on either the EC medium or on VRBA. In addition, one strain of *Yersinia enterocolitica* did not grow on the Compact Dry EC medium but did grow in VRBA. For Compact Dry EC, there were 7 strains giving atypical growth and 3 giving typical growth. For VRBA there were 5 strains giving atypical growth and 6 giving typical growth.

In the recent study (2018): Of the 10 exclusivity strains tested, three were detected by the alternate method and by the reference method these were *A.hydrophila* CRA 4111, *A.sobria* CRA 8390 and *S. fonticola* CRA 4613.

## Conclusion of the comparison studies

*For E.coli:* The results of the method comparison study showed that the Compact Dry EC provide equivalent results to the reference method ISO 16649-2:2001.

*For coliforms:* The results of the method comparison study showed that the Compact Dry EC provide equivalent results to the reference method ISO 4832:2006.

## INTERLABORATORY STUDY

### *E. coli*

The interlaboratory study was conducted in November 2007.

Number of laboratories: 9 [13 labs participated. 2 were excluded as the analysis were not performed on the agreed date and further 2 labs failed to test their samples for *E.coli* by the reference method.]


Samples: Pasteurised milk artificially contaminated with defined numbers of *E.coli*. The laboratories performed the analyses according to ISO 16649-2:2001 and Compact Dry EC method.

Table 6. Results (log cfu/g) – interlaboratory study of *E.coli*

Level	Reference method		Alternative method		Bias	Upper	Lower	Upper	Lower
	Median	S <sub>R</sub>	Alt method	S <sub>R</sub>		Level	Level	AL	AL
1	2.40	0.17	2.45	0.18	0.05	0.30	-0.04	0.50	-0.50
2	3.50	0.17	3.46	0.17	-0.04	0.20	-0.17	0.50	-0.50
3	4.38	0.34	4.50	0.20	0.12	0.40	0.01	0.50	-0.50

The results show that the bias is small and that the precision is satisfactory.

Figure 7. Accuracy Profile of the interlaboratory study for *E.coli*



## Coliforms

The interlaboratory study was conducted in November 2007.

Number of laboratories: 11


Samples: Pasteurised milk artificially contaminated with defined numbers of *E. coli*. The laboratories performed the analyses according to ISO 4832:2006 and Compact Dry EC method.

Table 7. Results (log cfu/g) – interlaboratory study of coliforms

Level	Referance method		Alternative method		Bias	Upper	Lower	Upper	Lower
	Median	S <sub>R</sub>	Alt method	S <sub>R</sub>		Level	Level	AL	AL
1	2.53	0.16	2.55	0.13	0.02	0.20	-0.09	0.50	-0.50
2	3.59	0.11	3.57	0.19	-0.02	0.25	-0.10	0.50	-0.50
3	4.48	0.075	4.59	0.15	0.11	0.32	0.00	0.50	-0.50

The results show that the bias is small and that the precision is satisfactory.

Figure 8 Accuracy Profile of the interlaboratory study for coliforms



## COLCLUSION

According to the comparison study and the interlaboratory study no substantial differences were found between the Compact Dry EC method and the reference method (ISO 16649-2:2001) for the enumeration of *Escherichia coli*, and no substantial differences were found between the Compact Dry EC method and the reference method (ISO 4832:2006) for the enumeration of coliforms.