

# Certificate of Analysis

Product CompactDry™ SL 1400plates ( Code 54058 )

Lot No 324309

Expiry Date 2025 . 2

	Specification	Results
Appearance	Blue purple. No particular matters.	Pass
Loss of drying	Less than 10.0%	7 . 0 %
Aseptic Test	Gamma irradiation over 16kGy	Pass
pH	5.5 ~ 6.3	6 . 0

## Evaluation of Performance Specification

- ① When following test strains are inoculated to incubate at 42°C for 20 - 24 hours, good growths and color developments should be observed.

<u>Test strain</u>	<u>Growth</u>	<u>Color development</u>
<i>Salmonella</i> Typhimurium ATCC 13311	+	yellow/green
<i>Salmonella</i> Enteritidis Sal- 1	+	yellow/green
<i>Salmonella</i> Habana Sal- 23-150	+	yellow
<i>Salmonella</i> Cubana Sal- FDA H <sub>2</sub> S-	+	yellow/green

- ② When following test strains are inoculated to incubate at 42°C for 20 - 24 hours, growths should be inhibited.

<u>Test strain</u>	
<i>Citrobacter freundii</i>	ATCC 8090
<i>Escherichia coli</i>	ATCC 8739
<i>Klebsiella pneumoniae</i>	ATCC 13883
<i>Proteus mirabilis</i>	ATCC 29906
<i>Staphylococcus aureus</i>	ATCC 6538

Judge: Acceptable quality.

Certificate issue date : 2023 . 9 . 28

Signed (QC Manager)



## Shimadzu Diagnostics Corporation

3-24-6 Ueno, TAITO-KU, TOKYO, 110-0005 JAPAN TEL: +81-3-5846-5611