

# Certificate of Analysis

Product CompactDry™ YMR 1400plates ( Code 54063 )

Lot No 141308

Expiry Date 2025 . 7

	Specification	Results
Appearance	Light yellow. No particular matters.	Pass
Loss of drying	Less than 10.0%	4 . 1 %
Aseptic Test	Gamma irradiation over 16kGy	Pass
pH	5.2 ~ 5.8	5 . 6

## Evaluation of Performance Specification

- ① When following test strains are inoculated to incubate at 25°C for 2 - 3 days, good growths and color developments should be observed.

<u>Test strain</u>	<u>Growth</u>	<u>Color development</u>
<i>Candida albicans</i> ATCC 10231	+	light yellow green
<i>Candida glabrata</i> JCM 3699	+	bluish green
<i>Aspergillus brasiliensis</i> ATCC 16404	+	blue


- ② When following test strains are inoculated to incubate at 25°C for 2 - 3 days, growths should be inhibited.

<u>Test strain</u>
<i>Bacillus subtilis</i> ATCC 6633
<i>Escherichia coli</i> ATCC 8739

Judge: Acceptable quality.

Certificate issue date : 2023 . 8 . 21

Signed (QC Manager)



**Shimadzu Diagnostics Corporation**

3-24-6 Ueno, TAITO-KU, TOKYO, 110-0005 JAPAN TEL: +81-3-5846-5611