

# Certificate of Analysis

Product AccuDia™ Trypto-Soya Broth (SCD Broth) (granule) ( Code 05630 )  
Lot No 113310  
Expiry Date 2026 . 9

	Specificati	Results
Appearance	Light brownish yellow granule	Pass
Loss of drying	Less than 6.0%	1 . 5 %
Dissolution	Clear light brownish yellow solution without any foreign insoluble matters	Pass
pH	7.1 ~ 7.5	7 . 3

## Evaluation of Performance Specification

When following test strains are inoculated to incubate, good growths should be observed.

<u>Test strain</u>		<u>incubation</u>
<i>Escherichia coli</i>	ATCC 8739	30°C for 20 - 24 hr
<i>Pseudomonas aeruginosa</i>	ATCC 9027	30°C for 20 - 24 hr
<i>Staphylococcus aureus</i>	ATCC 6538	30°C for 20 - 24 hr
<i>Bacillus subtilis</i>	ATCC 6633	25°C for 3 days
<i>Candida albicans</i>	ATCC 10231	25°C for 5 days
<i>Aspergillus brasiliensis</i>	ATCC 16404	25°C for 5 days

Judge: Acceptable quality.

Certificate issue date : 2023 . 10 . 6

Signed (QC Manager)



**Shimadzu Diagnostics Corporation**

3-24-6 Ueno, TAITO-KU, TOKYO, 110-0005 JAPAN TEL: +81-3-5846-5611