



## CERTIFICATE OF COMPLIANCE

### LRQA

hereby declares that the certification assessment has demonstrated that

### COMPACT DRY LM

for the detection of *Listeria monocytogenes*

Manufactured and supplied by:  
Shimadzu Diagnostics Corporation  
20th Floor Ueno Frontier Tower  
3-24-6 Ueno, Taito-ku, Tokyo, 110-8736  
JAPAN

has been validated and revealed to be at least equivalent to the reference method as demonstrated by the validation study report. The summary of the validation report is available on the MicroVal website: [www.microval.org](http://www.microval.org)

Reference method: ISO 11290-1:2017 Microbiology of the food chain — Horizontal method for the detection and enumeration of *Listeria monocytogenes* and of *Listeria spp.* Part 1: Detection method

Scope: broad range of foods covering meat and poultry products (RTE/RTRH), dairy products (raw and heat treated), fresh produce and fruit, seafood & fishery products, multicomponent foods, environmental samples

The validation and certification has been performed in accordance with EN ISO 16140-2:2016 and the MicroVal Rules and Certification Scheme version 9.1.

Certificate no.: 2020LR91b

First approval date: 11 March 2022

Surveillance date: 11 May 2023

Expiry date: 10 March 2026

ISSUED BY: LRQA Nederland B.V.  
Rotterdam, The Netherlands