

# Certificate of Analysis

Product CompactDry™ LS 1400plates ( Code 54060 )

Lot No 289403

Expiry Date 2025 . 8

	Specification	Results
Appearance	Light yellow. No particular matters.	Pass
Loss of drying	Less than 10.0%	4 . 4 %
Aseptic Test	Gamma irradiation over 16kGy	Pass
pH	7.0 ~ 7.4	7 . 3

## Evaluation of Performance Specification

- ① When following test strains are inoculated to incubate at 35°C for 22 - 26 hours, good growths and color developments should be observed.

<u>Test strain</u>		<u>Growth</u>	<u>Color development</u>
<i>Listeria monocytogenes</i>	JCM 7680	+	blue-green
<i>Listeria monocytogenes</i>	Lm-1	+	blue-green

- ② When following test strains are inoculated to incubate at 35°C for 22 - 26 hours, growths should be inhibited.

<u>Test strain</u>	
<i>Enterococcus faecalis</i>	ATCC 19433
<i>Bacillus subtilis</i>	ATCC 6633
<i>Escherichia coli</i>	ATCC 8739
<i>Pseudomonas aeruginosa</i>	ATCC 9027

Judge: Acceptable quality.

Certificate issue date : 2024 . 3 . 29

Signed (QC Manager)



**Shimadzu Diagnostics Corporation**

3-24-6 Ueno, TAITO-KU, TOKYO, 110-0005 JAPAN TEL: +81-3-5846-5611