

Certificate of Analysis

Product AccuDia™ GAM Agar, Modified (Code 05426)

Lot No 090310

Expiry Date 2026 . 9

| | Specification | Results |
|-----------------|--|---------|
| Appearance | Light brownish yellow fine powder | Pass |
| Loss of drying | Less than 6.0% | 3. 3 % |
| Dissolution | Clear light brown solution without any foreign insoluble matters | Pass |
| pH | 7.2 ~ 7.4 | 7. 2 |
| Strength of gel | 500~900 g/cm ² | 6 7 0 |

Evaluation of Performance Specification

When following test strains are inoculated to incubate at 37°C for 48 hours, good growths should be observed.

Test strain

Fusobacterium varium ATCC 8501

Clostridium perfringens ATCC 13123

Fusobacterium nucleatum JCM 6328

Eubacterium limosum ATCC 8486

Peptostreptococcus ascharolyticus ATCC 14963

Peptostreptococcus anaerobius ATCC 27337

Bacteroides fragilis ATCC 25285

Peptostreptococcus magnus ATCC 29328

Propionibacterium acnes ATCC 6919

Prevotella melaninogenica GAI 5492

Judge: Acceptable quality.

Certificate issue date : 2023 . 10 . 25

Signed (QC Manager)



Shimadzu Diagnostics Corporation

3-24-6 Ueno, TAITO-KU, TOKYO, 110-0005 JAPAN TEL: +81-3-5846-5611