

Certificate of Analysis

Product AccuDia™ Pyridoxine Assay Medium (Code 05815)
Lot No 092404
Expiry Date 2027 . 3

| | Specification | Results |
|----------------|--|---------|
| Appearance | Grayish white fine powder | Pass |
| Loss of drying | Less than 5.0% | 0 . 7 % |
| Dissolution | Clear light brownish yellow solution without any foreign insoluble matters | Pass |
| pH | 5.1 – 5.3 | 5 . 2 |

Evaluation of Performance Specification

When transmittance is measured after following test strain is incubate at 30°C for 16 - 24 hours and the calibration curve is made, it becomes a curve that can be mesured equally to the control medium.

Test strain
Saccharomyces carlsbergensis strain 4228 ATCC 9080

Judge: Acceptable quality.

Certificate issue date : 2024 . 4 . 18

Signed (QC Manager)



Shimadzu Diagnostics Corporation

3-24-6 Ueno, TAITO-KU, TOKYO, 110-0005 JAPAN TEL: +81-3-5846-5611