

Certificate of Analysis

Product AccuDia™ GAM Broth (Code 05422)

Lot No 128405

Expiry Date 2027 . 4

	Specification	Results
Appearance	Light brownish yellow fine powder	Pass
Loss of drying	Less than 6.0%	3 . 7 %
Dissolution	Clear light brown solution without any foreign insoluble matters	Pass
pH	7.0 ~ 7.2	7 . 1

Evaluation of Performance Specification

When following test strains are inoculated to incubate at 37°C for 48 hours, good growths should be observed.

Test strain

Fusobacterium varium ATCC 8501 *Bacteroides fragilis* ATCC 25285
Clostridium perfringens ATCC 13123 *Peptostreptococcus magnus* ATCC 29328
Fusobacterium nucleatum JCM 6328 *Propionibacterium acnes* ATCC 6919
Eubacterium limosum ATCC 8486 *Prevotella melaninogenica* GAI 5492
Peptostreptococcus asaccharolyticus ATCC 14963
Peptostreptococcus anaerobius ATCC 27337

Judge: Acceptable quality.

Certificate issue date : 2024 . 5 . 9

Signed (QC Manager)



Shimadzu Diagnostics Corporation

3-24-6 Ueno, TAITO-KU, TOKYO, 110-0005 JAPAN TEL: +81-3-5846-5611