

# Certificate of Analysis

Product AccuDia™ GAM Broth, Modified ( Code 05433 )

Lot No 029405

Expiry Date 2027 . 4

	Specification	Results
Appearance	Light yellow fine powder.	Pass
Loss of drying	Less than 6.0%	3 . 7 %
Dissolution	Clear light brownish yellow solution without any foreign insoluble matters	Pass
pH	7.2 ~ 7.4	7 . 3

## Evaluation of Performance Specification

When following test strains are inoculated to incubate at 37°C for 42 - 48 hours, good growths should be observed.

### Test strain

*Fusobacterium varium* ATCC 8501

*Clostridium perfringens* ATCC 13123

*Fusobacterium nucleatum* JCM 6328

*Eubacterium limosum* ATCC 8486

*Peptostreptococcus anaerobius* ATCC 27337

*Bacteroides thetaiotaomicron* ATCC 29741

*Bacteroides fragilis* ATCC 25285

*Peptostreptococcus asaccharolyticus* ATCC 14963

*Propionibacterium acnes* ATCC 6919

*Peptostreptococcus magnus* ATCC 29328

*Prevotella melaninogenica* GAI 5492

Judge: Acceptable quality.

Certificate issue date : 2024 . 5 . 14

Signed (QC Manager)



## Shimadzu Diagnostics Corporation

3-24-6 Ueno, TAITO-KU, TOKYO, 110-0005 JAPAN TEL: +81-3-5846-5611