

# Certificate of Analysis

Product AccuDia™ Biotin Assay Medium (Code 05818)

Lot No 088408

Expiry Date 2027 . 1

	Specification	Results
Appearance	Grayish white fine powder	Pass
Loss of drying	Less than 5.0%	1 . 2 %
Dissolution	Clear light brownish yellow solution without any foreign insoluble matters	Pass
pH	7.0 – 7.2	7 . 0

## Evaluation of Performance Specification

When transmittance is measured after following test strain is incubate at 37°C for 16 - 24 hours and the calibration curve is made, it becomes a curve that can be mesured equally to the control medium.

### Test strain

*Lactobacillus arabinosus* ATCC8014

Judge: Acceptable quality.

Certificate issue date : 2024 . 8 . 7

Signed (QC Manager)

*Takahisa Ueno*

**Shimadzu Diagnostics Corporation**

3-24-6 Ueno, TAITO-KU, TOKYO, 110-0005 JAPAN TEL: +81-3-5846-5611