

Certificate of Analysis

Product AccuDia™ Folic Acid Assay Medium (Code 05814)

Lot No 07740711

Expiry Date 2027 . 6

	Specification	Results
Appearance	Grayish white fine powder	Pass
Loss of drying	Less than 5.0%	0 . 8 %
Dissolution	Clear light brownish yellow solution without any foreign insoluble matters	Pass
pH	7.0 – 7.2	7 . 0

Evaluation of Performance Specification

When transmittance is measured after following test strain is incubate at 30°C for 16 - 24 hours and the calibration curve is made, it becomes a curve that can be mesured equally to the control medium.

Test strain
Streptococcus faecalis R ATCC 8043

Judge: Acceptable quality.

Certificate issue date : 2024 . 7 . 24

Signed (QC Manager)

Takahisa Ueno

Shimadzu Diagnostics Corporation

3-24-6 Ueno, TAITO-KU, TOKYO, 110-0005 JAPAN TEL: +81-3-5846-5611