

Certificate of Analysis

Product AccuDia™ TGC Medium without Indicator, Fluid (granule) (Code 05629)

Lot No 115409

Expiry Date 2027 . 8

	Specification	Results
Appearance	Light yellowish brown granule	Pass
Loss of drying	Less than 6.0%	1 . 1 %
Dissolution	Clear light yellowish brown solution without any foreign insoluble matters	Pass
pH	6.9 ~ 7.1	7 . 0

Evaluation of Performance Specification

When following test strains prepared to the level of concentration of 10 – 100 CFU are inoculated to incubate at 30°C for 3 days, good growths should be observed.

Test strain

Staphylococcus aureus ATCC 6538
Pseudomonas aeruginosa ATCC 9027
Clostridium sporogenes ATCC 11437

Judge: Acceptable quality.

Certificate issue date : 2024 . 9 . 20

Signed (QC Manager)



Shimadzu Diagnostics Corporation

3-24-6 Ueno, TAITO-KU, TOKYO, 110-0005 JAPAN TEL: +81-3-5846-5611