Certificate of Analysis

Product EC Blue 100 100mL x 80 tubes (Code 05593)

Lot № 234410

Expiry Date 2026.9

| | Specification | Results |
|----------------|--|---------|
| Appearance | Light yellow – dark yellow Granulated powder No particular matters | Pass |
| Loss of drying | Less than 6.0% | 1. 1 % |
| Dissolution | Clear light yellow solution without any foreign insoluble matters | Pass |
| pH | $6.9 \sim 7.3$ | 7.1 |
| Aseptic Test | No growth of any bacteria. | Pass |

Evaluation of Performance

Specification

When following test strains are inoculated to incubate at 37° C for 20 - 24 hours, growth and other developments should be comparable to those of control mediums.

| <u>Test strain</u> | Growth | <u>Color development</u> | <u>Fluorescence</u> |
|---------------------------------|--------|--------------------------|---------------------|
| Escherichia coli ATCC11775 | + | + - | + |
| Escherichia coli ATCC8739 | + | + | + |
| Enterobacter cloacae ATCC13047 | + | + | - |
| Citrobacter freundii ATCC8090 | + | + | - |
| Klebsiella pneumoniae ATCC13883 | + | + | - |
| Bacillus subtilis ATCC6633 | - | - | - |
| Staphylococcus aureus ATCC6538 | _ | - | - |

Judge: Acceptable quality.

Certificate issue date $:\!2024$. 11 . 8

Signed (QC Manager)

Jakahisa Douo

Shimadzu Diagnostics Corporation

3-24-6 Ueno, TAITO-KU, TOKYO, 110-0005 JAPAN TEL: +81-3-5846-5611