## **Certificate of Analysis**

Product EC Blue 100 100mL x 80 tubes (Code 05593)

Lot No 23441101

Expiry Date 2026.10

	Specification	Results
Appearance	Light yellow – dark yellow Granulated powder No particular matters	Pass
Loss of drying	Less than 6.0%	0.9%
Dissolution	Clear light yellow solution without any foreign insoluble matters	Pass
pH	$6.9 \sim 7.3$	7.1
Aseptic Test	No growth of any bacteria.	Pass

## **Evaluation of Performance**

Specification

When following test strains are inoculated to incubate at  $37^{\circ}$ C for 20 - 24 hours, growth and other developments should be comparable to those of control mediums.

<u>Test strain</u>	Growth	<u>Color development</u>	<u>Fluorescence</u>
Escherichia coli ATCC11775	+	+ -	+
Escherichia coli ATCC8739	+	+	+
Enterobacter cloacae ATCC13047	+	+	-
Citrobacter freundii ATCC8090	+	+	-
Klebsiella pneumoniae ATCC13883	+	+	-
Bacillus subtilis ATCC6633	_	-	-
Staphylococcus aureus ATCC6538	_	-	-

Judge: Acceptable quality.

Certificate issue date  $:\!2024$  . 12 . 13

Signed (QC Manager)

Jakahisa Douo

## Shimadzu Diagnostics Corporation

3-24-6 Ueno, TAITO-KU, TOKYO, 110-0005 JAPAN TEL: +81-3-5846-5611