

# Certificate of Analysis

Product        CompactDry™ BC 1400plates    ( Code 54068 )  
Lot №         101412  
Expiry Date    2026 . 8

	Specification	Results
Appearance	Light yellow No particular matters	Pass
Loss of drying	Less than 10.0%	4 . 3 %
Sterility	Gamma irradiation over 16kGy	Pass
pH	6.8 ~ 7.4	7 . 2

Evaluation of Performance  
Specification

① Productivity:

When following test strains are inoculated to incubate at 30℃ for 22 - 26 hours,  
good growths and color developments should be observed.

<u>Test strain</u>	<u>Growth</u>	<u>Color development</u>
<i>Bacillus cereus</i> ATCC 11778 (WDCM 00001)	+	blue
<i>Bacillus cereus</i> ATCC 19637	+	blue

② Selectivity and Specificity:

When following test strains are inoculated to incubate at 30℃ for 22 - 26 hours,  
growths should be inhibited, or color development different from that of *B.cereus*  
should be observed even when they grow.

<u>Test strain</u>	<u>Growth</u>	<u>Color development</u>
<i>Escherichia coli</i> ATCC 8739(WDCM 00012)	-	-
<i>Bacillus subtilis</i> ATCC 6633 (WDCM 00003)	- / +	- / white

Judge:    Acceptable quality.

Certificate issue date : 2024 . 12 . 27

Signed (QC Manager)

Takahisa Ueno

Shimadzu Diagnostics Corporation

3-24-6 Ueno, TAITO-KU, TOKYO, 110-0005 JAPAN    TEL: +81-3-5846-5611