

Certificate of Analysis

Product AccuDia™ BL Agar (Code 05430)

Lot No 075505

Expiry Date 2028 . 4

	Specification	Results
Appearance	Brownish yellow fine powder	Pass
Loss of drying	Less than 6.0%	3 . 4 %
Dissolution	Turbidity brownish yellow solution without any foreign insoluble matters	Pass
pH	7.1 ~ 7.3	7 . 1
Strength of gel	500~900 g/cm ²	5 1 0

Evaluation of Performance

Specification

When following test strains are inoculated to incubate at 37°C for 42 - 48 hours, good growths should be observed.

Test strain

Bacteroides fragilis ATCC 25285

Bacteroides vulgatus GAI 5460

Bifidobacterium longum Bif-3

Bifidobacterium bifidum Bif-4

Fusobacterium nucleatum JCM 6328

Peptostreptococcus aerogenes P-1

Peptostreptococcus anaerobius ATCC 27337

Propionibacterium acnes Prop-1

Judge: Acceptable quality.

Certificate issue date : 2025 . 5 . 22

Signed (QC Manager)

Takahisa Ueno

Shimadzu Diagnostics Corporation

3-24-6 Ueno, TAITO-KU, TOKYO, 110-0005 JAPAN TEL: +81-3-5846-5611