

Certificate of Analysis

Product CompactDry™ ECO 1400plates (Code 54064)

Lot No 056508

Expiry Date 2027 . 7

	Specification	Results
Appearance	Light yellow. No particular matters.	Pass
Loss of drying	Less than 10.0%	3 . 1 %
Aseptic Test	Gamma irradiation over 16kGy	Pass
pH	6.8 ~ 7.2	7 . 1

Evaluation of Performance

Specification

① When following test strains are inoculated to incubate at 35°C for 20-24hours,good growths and color developments should be observed.

<u>Test strain</u>	<u>Growth</u>	<u>Color development</u>
<i>Escherichia coli</i> ATCC8739	+	bluish green
<i>Escherichia coli</i> ATCC25922	+	bluish green
<i>Klebsiella oxytoca</i> ATCC13182	+	white
<i>Pseudomonas aeruginosa</i> ATCC9027	+	white

② When following test strains are inoculated to incubate at 35°C for 20-24hours,growths should be inhibited.

<u>Test strain</u>
<i>Bacillus subtilis</i> ATCC6633

Judge: Acceptable quality.

Certificate issue date : 2025 . 8 . 26

Signed (QC Manager)

Takahisa Ueno

Shimadzu Diagnostics Corporation

3-24-6 Ueno, TAITO-KU, TOKYO, 110-0005 JAPAN TEL: +81-3-5846-5611