

Certificate of Analysis

Product AccuDia™ B12 Inoculum Broth (Code 05803)

Lot № 069504

Expiry Date 2028 . 3

	Specification	Results
Appearance	Brownish yellow fine powder	Pass
Loss of drying	Less than 6.0%	2 . 9 %
Dissolution	Clear brownish yellow solution without any foreign insoluble matters	Pass
pH	6.8 – 7.0	6 . 8

Evaluation of Performance

Specification

When following test strain are inoculated to incubate at 37°C for 20-24 hours, good growths should be observed.

Test strain
Lactobacillus leichmannii ATCC 7830

Judge: Acceptable quality.

Certificate issue date : 2025 . 4 . 28

Signed (QC Manager)

Takahisa Ueno

Shimadzu Diagnostics Corporation

3-24-6 Ueno, TAITO-KU, TOKYO, 110-0005 JAPAN TEL: +81-3-5846-5611