

# Certificate of Analysis

Product AccuDia™ TGC Medium without Indicator, Fluid (granule) (Code 05629)

Lot No 117509

Expiry Date 2028 . 8

	Specification	Results
Appearance	Light yellowish brown granule	Pass
Loss of drying	Less than 6.0%	1 . 3 %
Dissolution	Clear light yellowish brown solution without any foreign insoluble matters	Pass
pH	6.9 ~ 7.1	6 . 9

## Evaluation of Performance Specification

When following test strains prepared to the level of concentration of 10 – 100 CFU are inoculated to incubate at 30°C for 3 days, good growths should be observed.

### Test strain

*Staphylococcus aureus* ATCC 6538  
*Pseudomonas aeruginosa* ATCC 9027  
*Clostridium sporogenes* ATCC 11437

Judge: Acceptable quality.

Certificate issue date : 2025 . 9 . 18

Signed (QC Manager)



**Shimadzu Diagnostics Corporation**

3-24-6 Ueno, TAITO-KU, TOKYO, 110-0005 JAPAN TEL: +81-3-5846-5611