

# Certificate of Analysis

Product CompactDry™ LM 1400plates ( Code 54067 )

Lot No 058512

Expiry Date 2027 . 11

	Specification	Results
Appearance	Light yellow. No particular matters.	Pass
Loss of drying	Less than 10.0%	4 . 0 %
Aseptic Test	Gamma irradiation over 16kGy	Pass
pH	7.0 ~ 7.6	7 . 4

## Evaluation of Performance Specification

When following test strains are inoculated to incubate at 37°C for 46 - 50 hours, good growths and color developments should be observed.

<u>Test strain</u>		<u>Growth</u>	<u>Color development</u>
<i>Listeria monocytogenes</i> ATCC 13932 (WDCM 00021)		+	red~red with blue surround
<i>Listeria monocytogenes</i> ATCC 35152 (WDCM 00109)		+	red~red with blue surround
<i>Listeria innocua</i> ATCC 33090 (WDCM 00017)		+	blue
<i>Escherichia coli</i> ATCC 8739 (WDCM 00012)		-	
<i>Enterococcus faecalis</i> ATCC 19433 (WDCM 00009)		-	

Judge: Acceptable quality.

Certificate issue date 2026 . 1 . 15

Signed (QC Manager)

反鋁 謙

Shimadzu Diagnostics Corporation