Certificate of Analysis

Product　CompactDry™ PA 1400plates （Code 54062）

Lot №　071512

Expiry Date　2027 . 6

	Specification	Results
Appearance	Light pink. No particular matters.	Pass
Loss of drying	Less than 10.0%	5 . 5 %
Aseptic Test	Gamma irradiation over 16kGy	Pass
pH	6.8 ～ 7.6	7 . 6

Evaluation of Performance

Specification

① When following test strains are inoculated to incubate at 35℃ for 45 - 51 hours, good growths and color developments should be observed.

Test strain	Growth	Color development
Pseudomonas aeruginosa ATCC 9027	＋	red
Pseudomonas aeruginosa ATCC 27853	＋	red
Pseudomonas aeruginosa ATCC 10145	＋	red

② When following test strains are inoculated to incubate at 35℃ for 20 - 24 hours, growths should be inhibited.

Test strain

Escherichia coli ATCC 8739

Bacillus subtilis ATCC 6633

Staphylococcus aureus ATCC 6538

Judge:　Acceptable quality.

Certificate issue date：2026 . 1 . 8

Signed (QC Manager)　石鍋　謙

Shimadzu Diagnostics Corporation

3-24-6 Ueno, TAITO-KU, TOKYO, 110-0005 JAPAN　TEL: +81-3-5846-5611