

Certificate of Analysis

Product AccuDia™ Clostridia Count Agar (Code 05409)

Lot No 099512

Expiry Date 2028 . 11

	Specification	Results
Appearance	Light brown fine powder	Pass
Loss of drying	Less than 6.0%	3 . 0 %
Dissolution	Clear light brown solution without any foreign insoluble matters	Pass
pH	7.5 – 7.7	7 . 7
Strength of gel	800 – 1100 g/cm ²	8 5 0

Evaluation of Performance Specification

When following test strains are inoculated to incubate at 35°C for 20 - 24 hours, following results should be observed respectively.

<u>Test strain</u>	<u>Growth</u>	<u>Colony</u>
<i>Clostridium perfringens</i> ATCC 13123	good	black colonies
<i>Clostridium sporogenes</i> ATCC 11437	good	black colonies
<i>Clostridium bifermentans</i> Cl-15	good	black colonies

Judge: Acceptable quality.

Certificate issue date : 2025 . 12 . 18

Signed (QC Manager)

反鍋 謙

Shimadzu Diagnostics Corporation

3-24-6 Ueno, TAITO-KU, TOKYO, 110-0005 JAPAN TEL: +81-3-5846-5611