

Certificate of Analysis

Product AccuDia™ Candida GE Agar (Code 05703)

Lot No 112507

Expiry Date 2028 . 6

	Specification	Results
Appearance	Light yellow powder	Pass
Loss of drying	Less than 6.0%	2 . 1 %
Dissolution	Clear light brownish yellow solution without any foreign insoluble matters	Pass
pH	5.9 –6.1	6 . 0
Strength of gel	500– 900 g/cm ²	5 6 0

Evaluation of Performance Specification

①When following test strains are inoculated to incubate at 37°C for 44 - 48 hours, good growths should be observed.

Test strain

Candida albicans

Candida krusei

②When following test strains are inoculated to incubate at 37°C for 44 - 48 hours, growths should be inhibited.

Test strain

Pseudomonas aeruginosa

Escherichia coli

Staphylococcus aureus

Judge: Acceptable quality.

Certificate issue date : 2025 . 7 . 23

Signed (QC Manager)

Takahisa Ueno

Shimadzu Diagnostics Corporation

3-24-6 Ueno, TAITO-KU, TOKYO, 110-0005 JAPAN TEL: +81-3-5846-5611