

Certificate of Analysis

Product CompactDry™ CF-L 1400plates (Code 54065)

Lot No 017512

Expiry Date 2027 . 11

	Specification	Results
Appearance	Light yellow. No particular matters.	Pass
Loss of drying	Less than 10.0%	5 . 1 %
Sterility	Gamma irradiation over 16kGy	Pass
pH	7.1 ~ 7.4	7 . 3

Evaluation of Performance

Specification

- ① When following test strains are inoculated to incubate at 35°C for 22-26hours, good growths and color developments should be observed.

<u>Test strain</u>	<u>Growth</u>	<u>Color development</u>
<i>Escherichia coli</i> ATCC8739	+	bluish green
<i>Klebsiella oxytoca</i> ATCC13182	+	bluish green
<i>Pseudomonas aeruginosa</i> ATCC9027	+	white

- ② When following test strains are inoculated to incubate at 35°C for 22-26hours, growths should be inhibited.

<u>Test strain</u>
<i>Bacillus subtilis</i> ATCC6633
<i>Staphylococcus aureus</i> ATCC6538

Judge: Acceptable quality.

Certificate issue date : 2025 . 12 . 17

Signed (QC Manager)

反鍋 謙

Shimadzu Diagnostics Corporation

3-24-6 Ueno, TAITO-KU, TOKYO, 110-0005 JAPAN TEL: +81-3-5846-5611