

Certificate of Analysis

Product CompactDry™ TCR 1400plates (Code 54069)

Lot No 075603

Expiry Date 2027. 8

	Specification	Results
Appearance	Light yellowish white. No particular matters.	Pass
Loss of drying	Less than 10.0%	4.4 %
Aseptic Test	Gamma irradiation over 16kGy	Pass
pH	6.8 ~ 7.2	7.0

Evaluation of Performance

Specification

When following test strains are inoculated to incubate at 35°C for 22 - 26 hours, good growths and color developments should be observed.

<u>Test strain</u>		<u>Growth</u>	<u>Color development</u>
<i>Bacillus subtilis</i>	ATCC 6633	+	red
<i>Escherichia coli</i>	ATCC 8739	+	red
<i>Escherichia coli</i>	ATCC 25922	+	red
<i>Staphylococcus aureus</i>	ATCC 25932	+	red

Judge: Acceptable quality.

Certificate issue date : 2026. 3. 30

Signed (QC Manager)

反鍋 謙

Shimadzu Diagnostics Corporation

3-24-6 Ueno, TAITO-KU, TOKYO, 110-0005 JAPAN TEL: +81-3-5846-5611