Certificate of Analysis

Product　CompactDry™ETC 1400plates （Code 54056）

Lot №　280605

Expiry Date　2027 . 10

	Specification	Results
Appearance	Light yellow. No particular matters.	Pass
Loss of drying	Less than 10.0%	４．５％
Aseptic Test	Gamma irradiation over 16kGy	Pass
pH	7.8 ～ 8.4	８．２

Evaluation of Performance

Specification

① When following test strains are inoculated to incubate at 37℃ for 20 - 24 hours, good growths and color developments should be observed.

Test strain	Growth	Color development
Enterococcus faecalis ATCC 19433	＋	blue
Enterococcus faecalis ATCC 51299	＋	blue

② When following test strains are inoculated to incubate at 35℃ for 20 - 24 hours, growths should be inhibited.

Test strain

Klebsiella oxytoca ATCC 13182

Bacillus subtilis ATCC 6633

Staphylococcus aureus ATCC 6538

Judge:　Acceptable quality.

Certificate issue date：2026 . 6 . 2

Signed (QC Manager)　石鍋　謙

Shimadzu Diagnostics Corporation

3-24-6 Ueno, TAITO-KU, TOKYO, 110-0005 JAPAN　TEL: +81-3-5846-5611