

Certificate of Analysis

Product AccuDia™ B12 Assay Medium (Set) (Code 05819)

Lot No 179607

Expiry Date 2029 . 6

	Specification	Results
Appearance	Grayish white fine powder	Pass
Loss of drying	Less than 5.0%	0 . 9 %
Dissolution	Clear light brownish yellow solution without any foreign insoluble matters	Pass
pH	6.1 – 6.3	6 . 2

Evaluation of Performance

Specification

When transmittance is measured after following test strain is incubate at 37°C for 16 - 24 hours and the calibration curve is made, it becomes a curve that can be mesured equally to the control medium.

Test strain

Lactobacillus leichmanii ATCC 7830

Judge: Acceptable quality.

Certificate issue date : 2026 . 7 . 17

Signed (QC Manager)

反 鍋 謙

Shimadzu Diagnostics Corporation

3-24-6 Ueno, TAITO-KU, TOKYO, 110-0005 JAPAN TEL: +81-3-5846-5611