

Certificate of Analysis

Product AccuDia™ GAM Broth, Modified (Code 05433)

Lot No 031605

Expiry Date 2029 . 4

	Specification	Results
Appearance	Light yellow fine powder.	Pass
Loss of drying	Less than 6.0%	3 . 0 %
Dissolution	Clear light brownish yellow solution without any foreign insoluble matters	Pass
pH	7.2 ~ 7.4	7 . 3

Evaluation of Performance Specification

When following test strains are inoculated to incubate at 37℃ for 42 - 48 hours, good growths should be observed.

Test strain

Fusobacterium varium ATCC 8501

Clostridium perfringens ATCC 13123

Fusobacterium nucleatum JCM 6328

Eubacterium limosum ATCC 8486

Peptostreptococcus anaerobius ATCC 27337

Bacteroides thetaiotaomicron ATCC 29741

Bacteroides fragilis ATCC 25285

Peptostreptococcus asaccharolyticus ATCC 14963

Propionibacterium acnes ATCC 6919

Peptostreptococcus magnus ATCC 29328

Prevotella melaninogenica GAI 5492

Judge: Acceptable quality.

Certificate issue date : 2026 . 5 . 11

Signed (QC Manager)

反鍋 謙

Shimadzu Diagnostics Corporation

3-24-6 Ueno, TAITO-KU, TOKYO, 110-0005 JAPAN TEL: +81-3-5846-5611