

Certificate of Analysis

Product AccuDia™ Pyridoxine Assay Medium (Code 05815)

Lot No 093506

Expiry Date 2027 . 12

	Specification	Results
Appearance	Grayish white fine powder	Pass
Loss of drying	Less than 5.0%	0 . 5 %
Dissolution	Clear light brownish yellow solution without any foreign insoluble matters	Pass
pH	5.1 – 5.3	5 . 2

Evaluation of Performance Specification

When transmittance is measured after following test strain is incubate at 30℃ for 16 - 24 hours and the calibration curve is made, it becomes a curve that can be mesured equally to the control medium.

Test strain

Saccharomyces carlsbergensis strain 4228 ATCC 9080

Judge: Acceptable quality.

Certificate issue date : 2025 . 6 . 27

Signed (QC Manager)

Takahisa Ueno

Shimadzu Diagnostics Corporation

3-24-6 Ueno, TAITO-KU, TOKYO, 110-0005 JAPAN TEL: +81-3-5846-5611